



Certificate of Acceptable Performance



THIS IS TO CERTIFY THAT:

**QUALITY CONTROL
DIASOURCE IMMUNOASSAYS SA
Lab 000825**

HAS SUCCESSFULLY COMPLETED THE EXTERNAL QUALITY ASSESSMENT FOR:

**EQAS Immunoassay (Monthly) Program
Cycle 13: December 2015 – December 2016**

FOR THE FOLLOWING TESTS:

11-Deoxycortisol
17-a-OH-Progesterone
25-OH-Vitamin D (Total)
Androstenedione
C-Peptide
CA 19-9
FSH
Growth Hormone
Insulin
TSH
Testosterone

Acceptable performance criteria: $\geq 75\%$ response rate with results within ± 3 Z-score



Bio-Rad Laboratories is fully accredited by the American Association for Laboratory Accreditation (A2LA) to the ISO/IEC 17043:2010 standard.

Azita Mokhtari

Azita Mokhtari, EQAS Program Manager

**Bio-Rad Laboratories
EQAS Program Enrollment**

9500 Jeronimo Road
Irvine, CA 92618-2017, USA
email: eqas@bio-rad.com